

CERTIFICATE OF CONFORMITY

Manufacturer: Ritchie Engineering Company, Inc.
10950 Hampshire Avenue South
Bloomington, MN 55438-2623
USA
Phone: 1-952-943-1300

Agent: Mr. Neil D. Stewart
Fairfield Technologies
Fairfield
St. Martins, UK
Phone: 44-1738-550-005

Declares that the product

Product Name: Yellow Jacket AccuProbe UV Leak Detector
Model Number(s): 69337

Conforms to the following Product Specifications:

EN 14624:2005 – Performances of Mobile Leak Detectors And Of Room Controllers Of Halogenated Refrigerants

Specification table in accordance with EN 14624 section 13:

Minimum sensitivity R134a fixed position	1gm/yr
Maximum sensitivity R134a fixed position	>50 gm/yr
Minimum sensitivity R134a moving	3 gm/yr
Maximum sensitivity R134a moving	>50 gm/yr
Minimum time for detection 1gram/yr	1 second
Zeroing time for 50 grams/yr exposure	Approx. 12 seconds
Minimum sensitivity in polluted ambiance	1 gram/yr fixed 3 grams/yr moving
Calibration Frequency	Check annually with calibrated leak source

Signed: 

Mr. Brian Flynn, Executive Vice President
Ritchie Engineering Company, Inc.

Date: 5 April 2011